



Low Asbestos Content Scheme Individual Results : Round 004 LACS Round 4

For Laboratory Number : 2069 Labor Dr. Schaeffner

From : 30/04/2018 To : 13/06/2018

Report Issued 29/06/2018

Report No. 4/733/400

Sample	Your Result	Assigned Result	Score
1	Amosite, Total Asbestos = 0.02%	Amosite, Total Asbestos = 0.09%	0 Z : -3.5

Health & Safety Laboratory
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team

Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 1

Page 1 of 1



Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK



Low Asbestos Content Scheme Individual Results : Round 005 LACS R5

For Laboratory Number : 2069 Labor Dr. Schaeffner

From : 03/09/2018 To : 12/10/2018

Report Issued 02/11/2018

Report No. 5/733/508

Sample	Your Result	Assigned Result	Score
1	Amosite, Total Asbestos = 0.06%	Amosite, Total Asbestos = 0.09%	0 Z : -1

Health & Safety Laboratory
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team

Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 1

Page 1 of 1



Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK



Low Asbestos Content Scheme Individual Results : Round 006 LACS R6

For Laboratory Number : 2069 Labor Dr. Schaeffner

From : 07/01/2019 To : 15/02/2019

Report Issued 06/03/2019

Report No. 6/733/623

Sample	Your Result	Assigned Result	Score
1	Crocidolite , Total Asbestos = 0.02%	Crocidolite , Total Asbestos = 0.03%	0 Z : -0.5

Health & Safety Laboratory
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team

Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 1

Page 1 of 1



Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK



Low Asbestos Content Scheme Individual Results : Round 007 LACS R7

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 29/04/2019 To : 07/06/2019

Report Issued 21/06/2019

Report No. 7/733/776

Sample	Your Result	Assigned Result	Score
1	Anthophyllite Total Asbestos = 0.04%	Anthophyllite Total Asbestos = 0.05%	0 Z : -0.5

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 008 LACS R8

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 02/09/2019 To : 11/10/2019

Report Issued 29/10/2019

Report No. 8/733/897

Sample	Your Result	Assigned Result	Score
1	Amosite Total Asbestos = 0.04%	Amosite Total Asbestos = 0.05%	0 Z : -0.5

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 009 LACS R9

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 06/01/2020 To : 14/02/2020

Report Issued 02/03/2020

Report No. 9/733/1026

Sample	Your Result	Assigned Result	Score
1	Amosite Total Asbestos = 0.04%	Amosite Total Asbestos = 0.06%	0 Z : -1

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 010 LACS R10

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 07/09/2020 To : 16/10/2020

Report Issued 06/11/2020

Report No. 10/733/1154

Sample	Your Result	Assigned Result	Score
1	Amosite Anthophyllite Total Asbestos = 0.04%	Amosite Anthophyllite Total Asbestos = 0.05%	0 Z : -0.5

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 011 LACS11

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 12/10/2020 To : 12/03/2021

Report Issued 26/03/2021

Report No. 11/733/1304

Sample	Your Result	Assigned Result	Score
1	Crocidolite Total Asbestos = 0.01%	Crocidolite Total Asbestos = 0.05%	0 Z : -3.33

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 012 LACS12

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 26/04/2021 To : 11/06/2021

Report Issued 29/06/2021

Report No. 12/733/1434

Sample	Your Result	Assigned Result	Score
1	Anthophyllite Total Asbestos = 0.05%	Anthophyllite Total Asbestos = 0.06%	0 Z : -0.5

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 013 LACS13

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 14/06/2021 To : 26/10/2021

Report Issued 05/11/2021

Report No. 15/733/1567

Sample	Your Result	Assigned Result	Score
1	Crocidolite Total Asbestos = 0.05%	Crocidolite Total Asbestos = 0.07%	0 Z : -1

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 014 LACS014

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 25/10/2021 To : 11/02/2022

Report Issued 28/02/2022

Report No. 17/733/1697

Sample	Your Result	Assigned Result	Score
1	Amosite Total Asbestos = 0.03%	Amosite Total Asbestos = 0.05%	0 Z : -1.18

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 018 LACS018

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 30/01/2023 To : 09/06/2023

Report Issued 22/06/2023

Report No. 25/733/2233

Sample	Your Result	Assigned Result	Score
1	Chrysotile	Chrysotile	0

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 019 LACS019

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 05/06/2023 To : 13/10/2023

Report Issued 30/10/2023

Report No. 26/733/2369

Sample	Your Result	Assigned Result	Score
1	Amosite	Amosite	0

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Low Asbestos Content Scheme Individual Results : Round 020 LACS020

For Laboratory Number : 2069 Labor Dr. Schaeffner GmbH

From : 16/10/2023 To : 26/02/2024

Report Issued 01/03/2024

Report No. 28/733/2501

Sample	Your Result	Assigned Result	Score
1	Amosite	Amosite	0

HSE
Melanie Clunas
LACS Scheme Co-Ordinator
Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1

