

SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

Laboratory No.	2069	Name Of Laboratory	Labor Dr. Schaeffner
Round No.	008A	Laboratory Representative	Dr. K.-H. Schaffner
		E-Mail Address	LABOR.Schaeffner@t-online.de
Report No.	0187	Report Issued	Nov 7 2018 11:20AM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
8A SEMS 1	1	0.00	0.00	69.00	0.00	0	A
	2						
	3						
8A SEMS 2	1	33.00	0.00	0.00	33.00	17.1	A
	2						
	3						
8A SEMS 3	1	51.00	0.00	0.00	51.00	29	A
	2						
	3						
8A SEMS 4	1	0.00	18.00	0.00	18.00	16.4	A
	2						
	3						

Total Number In Each Band

-C	-B	A	B	C
0	0	4	0	0

Details of performance assesement are given in
'SEMS Information For Participants'.

Number Of Valid Results **4**

Results Within Band A **4** **100%**

Results Within Band A + B **4** **100%**

Melanie Clunas
SEM Scheme Co-Ordinator
Proficiency Testing Team



Health & Safety Laboratory
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK
Report No. : 0187



SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

Laboratory No.	2069	Name Of Laboratory	Labor Dr. Schaeffner GmbH
Round No.	009A	Laboratory Representative	Martin Schmittner
		E-Mail Address	LABOR.Schaeffner@t-online.de
Report No.	0381	Report Issued	Nov 6 2019 8:29AM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
Sample 1	1	16.00	0.00	0.00	16.00	12	A
	2						
	3						
Sample 2	1	44.00	0.00	0.00	44.00	49	A
	2						
	3						
Sample 3	1	29.00	0.00	0.00	29.00	26.5	A
	2						
	3						
Sample 4	1	88.00	0.00	0.00	88.00	65.3	A
	2						
	3						

Total Number In Each Band

-C -B A B C

0 0 4 0 0

Details of performance assesement are given in
'SEMS Information For Participants'.

Number Of Valid Results

4

Results Within Band A

4

100%

Results Within Band A + B

4

100%

Melanie Clunas
SEM Scheme Co-Ordinator
Proficiency Testing Team

HSE, Science Research Centre
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK
Report No. : 0381



SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

This report is a re-issue of 0505 and supercedes all previous issues.

RE-ISSUE REASON : Calculation Issue

Sample 1 - the median was incorrect. Sample 3 - the band information for an A score was incorrect

Laboratory No.	2069	Name Of Laboratory	Labor Dr. Schaeffner GmbH
Round No.	010A	Laboratory Representative	Martin Schmittner
		E-Mail Address	LABOR.Schaeffner@t-online.de
Report No.	0556	Report Issued	Apr 26 2021 4:24PM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
Sample 1	1	54.00		1.00	54.00	40.8	A
	2						
	3						
Sample 2	1	34.00		1.00	34.00	28	A
	2						
	3						
Sample 3	1	97.00		1.00	97.00	74.7	A
	2						
	3						
Sample 4	1	41.00		1.00	41.00	30	A
	2						
	3						

Total Number In Each Band

-C -B A B C

0 0 4 0 0

Details of performance assesement are given in
'SEMS Information For Participants'.

Number Of Valid Results

4

Results Within Band A

4

100%

Results Within Band A + B

4

100%

Melanie Clunas
SEM Scheme Co-Ordinator
Proficiency Testing Team

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK
Report No. : 0556



SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

Laboratory No.	2069	Name Of Laboratory	Labor Dr. Schaeffner GmbH
Round No.	011C	Laboratory Representative	Martin Schmittner
		E-Mail Address	LABOR.Schaeffner@t-online.de
Report No.	0674	Report Issued	Nov 3 2021 12:45PM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
1	1	99.00			99.00	72	A
	2						
	3						
2	1	48.00			48.00	44.2	A
	2						
	3						
3	1			40.00	0.00	0	A
	2						
	3						
4	1		29.00		29.00	20.6	A
	2						
	3						

Total Number In Each Band

-C	-B	A	B	C
0	0	4	0	0

Details of performance assesement are given in
'SEMS Information For Participants'.

Number Of Valid Results

4

Results Within Band A

4

100%

Results Within Band A + B

4

100%

Melanie Clunas
SEM Scheme Co-Ordinator
Proficiency Testing Team

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK
Report No. : 0674



SCANNING ELECTRON MICROSCOPY SCHEME (SEMS)

Laboratory No.	2069	Name Of Laboratory	Labor Dr. Schaeffner GmbH
Round No.	012A	Laboratory Representative	Martin Schmittner
		E-Mail Address	LABOR.Schaeffner@t-online.de
Report No.	0726	Report Issued	Nov 2 2022 2:57PM

SAMPLE	No	Amphibole Density	Chrysotile Density	Inorganic Density	Total Asbestos	Median	Band
1	1	0.00	29.50	0.00	29.50	25.2	A
	2						
	3						
2	1	9.00	0.00	0.00	9.00	11	A
	2						
	3						
3	1	89.00	0.00	3.00	89.00	80.6	A
	2						
	3						
4	1	0.00	0.00	96.50	0.00	0	A
	2						
	3						

Total Number In Each Band

-C	-B	A	B	C
0	0	4	0	0

Details of performance assesement are given in
'SEMS Information For Participants'.

Number Of Valid Results

4

Results Within Band A

4

100%

Results Within Band A + B

4

100%

Melanie Clunas
SEM Scheme Co-Ordinator
Proficiency Testing Team

HSE, Science & Research Centre
Harpur Hill, Buxton, Derbyshire SK17 9JN - UK
Report No. : 0726